

CERTIFICATE OF ANALYSIS

For

Lithium Regeneration Buffer 6 NC

Part Number: 80-6003-06

Batch (Control) Number: P27751NC

Quantity: 1 Litre

Expiry Date: 04 May 2024

The above product gave the following results on analysis:

Conductivity = 51.1 mScm⁻¹ @ 20.0°C

Conductivity Specification = 49.6 - 55.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

David Thomas

Date of Issue: 21 May 2021

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.