

CERTIFICATE OF ANALYSIS

For

Polyamine Buffer

Part Number: 80-2037-74

Batch (Control) Number: P27752

Quantity: 2 Litres

Expiry Date: 07 May 2024

The above product gave the following results on analysis:

pH = 5.615 @ 20.0°C

pH Specification = 5.58 - 5.62 @ 20.0°C

Conductivity = 91.0 mScm⁻¹ @ 20.0°C

Conductivity Specification = 83.6 - 94.5 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 10 May 2021

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.