

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 3

Part Number: 80-2099-83

Batch (Control) Number: R27799

Quantity: 2 Litres

Expiry Date: 11 April 2025

The above product gave the following results on analysis:

pH = 3.134 @ 20.0°C

pH Specification = 3.13 - 3.17 @ 20.0°C

Conductivity = 34.7 mScm⁻¹ @ 20.0°C

Conductivity Specification = 33.4 - 39.7 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 19 April 2022

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.