

# CERTIFICATE OF ANALYSIS

For

## Polyamine Buffer

Part Number: 80-2037-74

Batch (Control) Number: R27820

Quantity: 2 Litres

Expiry Date: 03 October 2025

The above product gave the following results on analysis:

pH = 5.593 @ 20.0°C

pH Specification = 5.58 - 5.62 @ 20.0°C

Conductivity = 93.8 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 83.6 - 94.5 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:

Date of Issue: 10 October 2022

Neil Booth

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.