

CERTIFICATE OF ANALYSIS

For

Sodium Loading Buffer

Part Number: 80-2037-67

Batch (Control) Number: T27833

Quantity: 2 Litres

Expiry Date: 17 January 2026

The above product gave the following results on analysis:

pH = 2.187 @ 20.0°C

pH Specification = 2.18 - 2.22 @ 20.0°C

Conductivity = 18.63 mScm⁻¹ @ 20.0°C

Conductivity Specification = 17.7 - 20.0 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 24 January 2023

Neil Booth

| | | | |
|---|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England. | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.