

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 1 NC

Part Number: 80-6003-01

Batch (Control) Number: T27836NC

Quantity: 2 Litres

Expiry Date: 14 February 2026

The above product gave the following results on analysis:

pH = 2.796 @ 20.0°C pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 14.78 mScm-1 @ 20.0°C Conductivity Specification = 14.4 - 16.0 mScm-1 @ 20.0°C

Chemical QC Signature: Date of Issue: 20 February 2023

David Thomas

Postal Address Telephone

+44 (0)1223 423723

Telefax +44 (0)1223 420164

support@biochrom.co.uk

Email:

Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach

Cambridge

Biochrom Ltd

CB25 9PE, England. Web Site: http://www.biochrom.co.uk

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.

Certificate of Analysis v1.1 - 29 September 2020.