

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 4 NC

Part Number: 80-6003-04

Batch (Control) Number: T27843NC

Quantity: 2 Litres

Expiry Date: 14 March 2026

The above product gave the following results on analysis:

pH = 3.489 @ 20.0°C

pH Specification = 3.48 - 3.52 @ 20.0°C

Conductivity = 56.7 mScm⁻¹ @ 20.0°C

Conductivity Specification = 54.9 - 58.2 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

David Thomas

Date of Issue: 17 March 2023

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.